

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032824\
 Data File : VD078780.D
 Acq On : 28 Mar 2024 17:52
 Operator : RP/MD
 Sample : P1879-05
 Misc : 4.17G/5ml/MSVOA_D/SOIL/B
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EB-03(0-2)

Quant Time: Mar 29 01:37:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 15 05:38:03 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	141447	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	271247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	235076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	91579	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	66774	50.373	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.740%
35) Dibromofluoromethane	7.810	113	83356	47.730	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.460%
50) Toluene-d8	10.269	98	328475	51.142	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.280%
62) 4-Bromofluorobenzene	12.575	95	75676	40.274	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	80.540%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD032824\
Data File : VD078780.D
Acq On : 28 Mar 2024 17:52
Operator : RP/MD
Sample : P1879-05
Misc : 4.17G/5ml/MSVOA_D/SOIL/B
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EB-03(0-2)

Quant Time: Mar 29 01:37:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031424S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 15 05:38:03 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. 0.000 min
Lab File: VD078780.D
Acq: 28 Mar 2024 17:52




Instrument :
MSVOA_D
ClientSampleId :
EB-03(0-2)


Tgt Ion:168 Resp: 141447
Ion Ratio Lower Upper
168 100
99 50.3 38.8 58.2



#33
1,2-Dichloroethane-d4
Concen: 50.373 ug/l
RT: 8.234 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VD078780.D
Acq: 28 Mar 2024 17:52

Tgt Ion: 65 Resp: 66774
Ion Ratio Lower Upper
65 100
67 53.4 0.0 110.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 1

Delta R.T. 0.006 min

Lab File: VD078780.D

Acq: 28 Mar 2024 17:52

Instrument :

MSVOA_D

ClientSampleId :

EB-03(0-2)





Tgt Ion:114 Resp: 271247

Ion Ratio Lower Upper

114 100

63 16.3 0.0 33.4

88 12.5 0.0 24.6



#35

Dibromofluoromethane

Concen: 47.730 ug/l

RT: 7.810 min Scan# 1058

Delta R.T. 0.006 min

Lab File: VD078780.D

Acq: 28 Mar 2024 17:52




Tgt Ion:113 Resp: 83356


Ion Ratio Lower Upper

113 100

111 104.2 81.9 122.9

192 18.7 15.4 23.0









#50
Toluene-d8
Concen: 51.142 ug/l
RT: 10.269 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VD078780.D
Acq: 28 Mar 2024 17:52




Instrument :
MSVOA_D
ClientSampleId :
EB-03(0-2)


Tgt Ion: 98 Resp: 328475
Ion Ratio Lower Upper
98 100
100 64.3 50.5 75.7



#62
4-Bromofluorobenzene
Concen: 40.274 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. 0.006 min
Lab File: VD078780.D
Acq: 28 Mar 2024 17:52

Tgt Ion: 95 Resp: 75676
Ion Ratio Lower Upper
95 100
174 95.8 0.0 195.2
176 94.0 0.0 188.2





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.581 min Scan# 1

Delta R.T. 0.000 min

Lab File: VD078780.D

Acq: 28 Mar 2024 17:52

Instrument :

MSVOA_D

ClientSampleId :

EB-03(0-2)





Tgt Ion:117 Resp: 235076

Ion Ratio Lower Upper

117 100

82 50.4 41.4 62.2

119 30.6 26.5 39.7



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.516 min Scan# 2028

Delta R.T. -0.000 min

Lab File: VD078780.D

Acq: 28 Mar 2024 17:52

Tgt Ion:152 Resp: 91579

Ion Ratio Lower Upper

152 100

115 52.7 27.1 81.2

150 155.4 0.0 350.2

