

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD032919\
 Data File : VD061437.D
 Acq On : 29 Mar 2019 18:42
 Operator : VA/AP
 Sample : K1698-03
 Misc : 8.17g/5ml/MSVOA D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-05-032819

Quant Time: Mar 30 00:17:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	38078	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.22	114	73041	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.37	117	64811	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.51	152	25406	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	31981	98.32	ug/l	0.02
Spiked Amount 50.000			Recovery	=	196.64%	
35) Dibromofluoromethane	6.93	113	36970	66.23	ug/l	0.03
Spiked Amount 50.000			Recovery	=	132.46%	
50) Toluene-d8	10.41	98	64774	48.38	ug/l	0.02
Spiked Amount 50.000			Recovery	=	96.76%	
62) 4-Bromofluorobenzene	13.56	95	32334	62.98	ug/l	0.01
Spiked Amount 50.000			Recovery	=	125.96%	
Target Compounds						
16) Acetone	3.22	43	31313	374.477	ug/l	91
20) Methylene Chloride	3.83	84	4257	7.454	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD032919\
Data File : VD061437.D
Acq On : 29 Mar 2019 18:42
Operator : VA/AP
Sample : K1698-03
Misc : 8.17g/5ml/MSVOA D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-05-032819

Quant Time: Mar 30 00:17:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

