

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033022\
 Data File : VD072476.D
 Acq On : 30 Mar 2022 13:52
 Operator : VA/SY
 Sample : N2103-07RE
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RBR-200049RE

Quant Time: Mar 31 03:08:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	53644	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	94223	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	121214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	60935	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	37996	58.552	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.100%
35) Dibromofluoromethane	7.914	113	37271	50.900	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.800%
50) Toluene-d8	10.338	98	116344	45.709	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	91.420%
62) 4-Bromofluorobenzene	12.626	95	57906	58.252	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033022\
Data File : VD072476.D
Acq On : 30 Mar 2022 13:52
Operator : VA/SY
Sample : N2103-07RE
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RBR-200049RE

Quant Time: Mar 31 03:08:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

