

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033022\
 Data File : VD072483.D
 Acq On : 30 Mar 2022 17:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 03:10:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.397	0.415	-4.5	85	0.00
3 P	Chloromethane	0.367	0.386	-5.2	81	0.00
4 C	Vinyl Chloride	0.385	0.428	-11.2#	84	0.00
5 T	Bromomethane	0.283	0.294	-3.9	84	0.00
6 T	Chloroethane	0.270	0.295	-9.3	85	0.00
7 T	Trichlorofluoromethane	0.861	0.925	-7.4	85	0.00
8 T	Diethyl Ether	0.256	0.263	-2.7	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.574	0.600	-4.5	86	0.00
10 T	Methyl Iodide	0.457	0.531	-16.2	84	0.00
11 T	Tert butyl alcohol	0.032	0.030	6.3	84	0.00
12 CM	1,1-Dichloroethene	0.449	0.475	-5.8#	82	0.00
13 T	Acrolein	0.014	0.007	50.0#	71	0.00
14 T	Allyl chloride	0.721	0.738	-2.4	79	0.00
15 T	Acrylonitrile	0.128	0.125	2.3	84	0.00
16 T	Acetone	0.113	0.115	-1.8	87	0.00
17 T	Carbon Disulfide	0.795	1.006	-26.5#	83	0.00
18 T	Methyl Acetate	0.318	0.270	15.1	84	0.00
19 T	Methyl tert-butyl Ether	1.311	1.325	-1.1	84	0.00
20 T	Methylene Chloride	0.671	0.637	5.1	86	0.00
21 T	trans-1,2-Dichloroethene	0.502	0.569	-13.3	85	0.00
22 T	Diisopropyl ether	1.804	1.850	-2.5	84	0.00
23 T	Vinyl Acetate	0.871	0.953	-9.4	83	0.00
24 P	1,1-Dichloroethane	1.087	1.100	-1.2	84	-0.01
25 T	2-Butanone	0.156	0.157	-0.6	84	0.00
26 T	2,2-Dichloropropane	0.966	0.983	-1.8	85	0.00
27 T	cis-1,2-Dichloroethene	0.682	0.712	-4.4	85	0.00
28 T	Bromochloromethane	0.466	0.467	-0.2	88	0.00
29 T	Tetrahydrofuran	0.094	0.091	3.2	80	0.00
30 C	Chloroform	1.223	1.248	-2.0#	87	0.00
31 T	Cyclohexane	0.791	0.849	-7.3	84	0.00
32 T	1,1,1-Trichloroethane	1.037	1.095	-5.6	87	0.00
33 S	1,2-Dichloroethane-d4	0.605	0.550	9.1	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.389	0.364	6.4	84	0.00
36 T	1,1-Dichloropropene	0.455	0.510	-12.1	85	0.00
37 T	Ethyl Acetate	0.206	0.201	2.4	81	0.00
38 T	Carbon Tetrachloride	0.535	0.584	-9.2	86	0.00
39 T	Methylcyclohexane	0.498	0.576	-15.7	84	0.00
40 TM	Benzene	1.357	1.499	-10.5	86	0.00
41 T	Methacrylonitrile	0.112	0.105	6.3	84	0.00
42 TM	1,2-Dichloroethane	0.415	0.445	-7.2	87	0.00
43 T	Isopropyl Acetate	0.395	0.400	-1.3	83	0.00
44 TM	Trichloroethene	0.385	0.415	-7.8	84	0.00
45 C	1,2-Dichloropropane	0.377	0.401	-6.4#	87	0.00
46 T	Dibromomethane	0.198	0.214	-8.1	87	0.00
47 T	Bromodichloromethane	0.569	0.603	-6.0	90	0.00
48 T	Methyl methacrylate	0.192	0.204	-6.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033022\
 Data File : VD072483.D
 Acq On : 30 Mar 2022 17:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 03:10:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	90	0.00
50 S	Toluene-d8	1.351	1.256	7.0	83	0.00
51 T	4-Methyl-2-Pentanone	0.212	0.216	-1.9	84	0.00
52 CM	Toluene	0.902	1.009	-11.9#	87	0.00
53 T	t-1,3-Dichloropropene	0.496	0.535	-7.9	86	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.628	-7.2	87	0.00
55 T	1,1,2-Trichloroethane	0.300	0.315	-5.0	88	0.00
56 T	Ethyl methacrylate	0.339	0.366	-8.0	84	0.00
57 T	1,3-Dichloropropane	0.496	0.530	-6.9	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.148	0.149	-0.7	83	0.00
59 T	2-Hexanone	0.144	0.156	-8.3	87	0.00
60 T	Dibromochloromethane	0.387	0.409	-5.7	89	0.00
61 T	1,2-Dibromoethane	0.267	0.289	-8.2	89	0.00
62 S	4-Bromofluorobenzene	0.528	0.492	6.8	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.327	0.369	-12.8	90	0.00
65 PM	Chlorobenzene	1.118	1.212	-8.4	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.447	0.465	-4.0	87	0.00
67 C	Ethyl Benzene	1.951	2.176	-11.5#	88	0.00
68 T	m/p-Xylenes	0.740	0.839	-13.4	89	0.00
69 T	o-Xylene	0.714	0.794	-11.2	88	0.00
70 T	Styrene	1.258	1.404	-11.6	88	0.00
71 P	Bromoform	0.237	0.246	-3.8	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	4.038	4.226	-4.7	87	0.00
74 T	N-amyl acetate	0.847	0.834	1.5	84	0.00
75 P	1,1,2,2-Tetrachloroethane	0.765	0.747	2.4	91	0.00
76 T	1,2,3-Trichloropropane	0.545	0.502	7.9	74	0.00
77 T	Bromobenzene	0.905	0.936	-3.4	89	0.00
78 T	n-propylbenzene	4.871	5.163	-6.0	88	0.00
79 T	2-Chlorotoluene	2.849	2.954	-3.7	88	0.00
80 T	1,3,5-Trimethylbenzene	3.458	3.612	-4.5	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.222	0.221	0.5	87	0.00
82 T	4-Chlorotoluene	2.967	3.072	-3.5	90	0.00
83 T	tert-Butylbenzene	2.994	3.097	-3.4	89	0.00
84 T	1,2,4-Trimethylbenzene	3.382	3.561	-5.3	91	0.00
85 T	sec-Butylbenzene	4.510	4.736	-5.0	89	0.00
86 T	p-Isopropyltoluene	3.686	3.900	-5.8	90	0.00
87 T	1,3-Dichlorobenzene	1.875	1.928	-2.8	92	0.00
88 T	1,4-Dichlorobenzene	1.890	1.924	-1.8	91	0.00
89 T	n-Butylbenzene	3.512	3.727	-6.1	90	0.00
90 T	Hexachloroethane	0.755	0.764	-1.2	91	0.00
91 T	1,2-Dichlorobenzene	1.654	1.694	-2.4	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.113	0.111	1.8	90	0.00
93 T	1,2,4-Trichlorobenzene	1.025	1.071	-4.5	91	0.00
94 T	Hexachlorobutadiene	0.574	0.599	-4.4	94	0.00
95 T	Naphthalene	1.992	1.977	0.8	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033022\
Data File : VD072483.D
Acq On : 30 Mar 2022 17:15
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Mar 31 03:10:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.895	0.939	-4.9	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6