

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD033119\
 Data File : VD061459.D
 Acq On : 31 Mar 2019 20:55
 Operator : VA/AP
 Sample : K2131-04RE
 Misc : 7.01g/5ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-GF-02-035-BRE

Quant Time: Apr 01 06:56:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	534541	50.00	ug/l	0.05
34) 1,4-Difluorobenzene	8.24	114	918005	50.00	ug/l	0.05
63) Chlorobenzene-d5	12.38	117	686822	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	14.53	152	298458	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	219473	48.06	ug/l	0.04
Spiked Amount 50.000			Recovery	=	96.12%	
35) Dibromofluoromethane	6.95	113	307216	43.79	ug/l	0.05
Spiked Amount 50.000			Recovery	=	87.58%	
50) Toluene-d8	10.42	98	761778	45.27	ug/l	0.04
Spiked Amount 50.000			Recovery	=	90.54%	
62) 4-Bromofluorobenzene	13.56	95	290980	45.10	ug/l	0.02
Spiked Amount 50.000			Recovery	=	90.20%	
Target Compounds						
16) Acetone	3.24	43	18497	15.758	ug/l	99
20) Methylene Chloride	3.84	84	29994	2.747	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD033119\
Data File : VD061459.D
Acq On : 31 Mar 2019 20:55
Operator : VA/AP
Sample : K2131-04RE
Misc : 7.01g/5ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-GF-02-035-BRE

Quant Time: Apr 01 06:56:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

