


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD033120\
 Data File : VD065557.D
 Acq On : 31 Mar 2020 10:37
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032620S.M
 Title : SW846 8260
 Last Update : Fri Mar 27 01:47:48 2020



AutoFind: Scans 1822, 1823, 1824; Background Corrected with Scan 1813

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	54952	PASS
75	95	30	60	47.9	147733	PASS
95	95	100	100	100.0	308117	PASS
96	95	5	9	6.7	20682	PASS
173	174	0.00	2	1.3	3699	PASS
174	95	50	100	94.1	290026	PASS
175	174	5	9	7.3	21053	PASS
176	174	95	101	97.0	281429	PASS
177	176	5	9	6.7	18800	PASS