

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
 Data File : VD068736.D
 Acq On : 31 Mar 2021 14:17
 Operator : VA/SY
 Sample : M1874-03
 Misc : 6.50G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB18(39-40)

Quant Time: Apr 01 00:44:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	334119	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	604091	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	632824	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	307578	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	181697	53.93	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 107.86%		
35) Dibromofluoromethane	7.920	113	195642	52.05	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 104.10%		
50) Toluene-d8	10.338	98	775594	53.52	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 107.04%		
62) 4-Bromofluorobenzene	12.632	95	289481	61.61	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 123.22%		
Target Compounds						Qvalue
16) Acetone	4.156	43	4016	9.83	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
Data File : VD068736.D
Acq On : 31 Mar 2021 14:17
Operator : VA/SY
Sample : M1874-03
Misc : 6.50G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB18(39-40)

Quant Time: Apr 01 00:44:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

