

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
 Data File : VD068738.D
 Acq On : 31 Mar 2021 15:13
 Operator : VA/SY
 Sample : M1874-05
 Misc : 5.88G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB16(26-27)

Quant Time: Apr 01 00:44:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	331295	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	602973	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	620049	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	299521	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	172943	51.77	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.54%	
35) Dibromofluoromethane	7.914	113	188822	50.33	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.66%	
50) Toluene-d8	10.338	98	760904	52.61	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.22%	
62) 4-Bromofluorobenzene	12.632	95	282469	60.23	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 120.46%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
Data File : VD068738.D
Acq On : 31 Mar 2021 15:13
Operator : VA/SY
Sample : M1874-05
Misc : 5.88G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB16(26-27)

Quant Time: Apr 01 00:44:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

