

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
 Data File : VD068739.D
 Acq On : 31 Mar 2021 15:42
 Operator : VA/SY
 Sample : M1874-06
 Misc : 5.63G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FS-SB16(45-46)

Quant Time: Apr 01 00:45:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	35367	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	58414	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	71490	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	36012	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	26193	73.45	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.90%
35) Dibromofluoromethane	7.914	113	20927	57.58	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	115.16%
50) Toluene-d8	10.344	98	76906	54.88	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.76%
62) 4-Bromofluorobenzene	12.626	95	28406	62.53	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.06%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
Data File : VD068739.D
Acq On : 31 Mar 2021 15:42
Operator : VA/SY
Sample : M1874-06
Misc : 5.63G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FS-SB16(45-46)

Quant Time: Apr 01 00:45:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

