

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
 Data File : VD068730.D
 Acq On : 31 Mar 2021 11:01
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 01 00:41:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	94	0.00
2 T	Dichlorodifluoromethane	0.391	0.380	2.8	95	0.00
3 P	Chloromethane	0.482	0.436	9.5	89	0.00
4 C	Vinyl Chloride	0.569	0.571	-0.4#	97	0.00
5 T	Bromomethane	0.421	0.378	10.2	93	0.00
6 T	Chloroethane	0.373	0.362	2.9	94	0.00
7 T	Trichlorofluoromethane	0.678	0.710	-4.7	101	0.00
8 T	Diethyl Ether	0.188	0.182	3.2	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.406	0.436	-7.4	103	0.00
10 T	Methyl Iodide	0.419	0.423	-1.0	86	0.00
11 T	Tert butyl alcohol	0.023	0.022	4.3	91	0.00
12 CM	1,1-Dichloroethene	0.395	0.403	-2.0#	95	0.00
13 T	Acrolein	0.021	0.018	14.3	96	-0.01
14 T	Allyl chloride	0.480	0.455	5.2	90	0.00
15 T	Acrylonitrile	0.091	0.087	4.4	88	0.00
16 T	Acetone	0.061	0.070	-14.8	112	0.00
17 T	Carbon Disulfide	1.269	1.247	1.7	92	0.00
18 T	Methyl Acetate	0.168	0.168	0.0	92	0.00
19 T	Methyl tert-butyl Ether	0.995	0.941	5.4	88	0.00
20 T	Methylene Chloride	0.666	0.510	23.4	94	0.00
21 T	trans-1,2-Dichloroethene	0.549	0.530	3.5	91	0.00
22 T	Diisopropyl ether	1.220	1.132	7.2	86	0.00
23 T	Vinyl Acetate	0.648	0.602	7.1	85	0.00
24 P	1,1-Dichloroethane	0.895	0.855	4.5	91	0.00
25 T	2-Butanone	0.101	0.090	10.9	82	0.00
26 T	2,2-Dichloropropane	0.796	0.707	11.2	84	0.00
27 T	cis-1,2-Dichloroethene	0.612	0.517	15.5	79	0.00
28 T	Bromochloromethane	0.354	0.322	9.0	83	0.00
29 T	Tetrahydrofuran	0.064	0.057	10.9	81	0.00
30 C	Chloroform	0.973	0.836	14.1#	82	0.00
31 T	Cyclohexane	0.813	0.752	7.5	91	0.00
32 T	1,1,1-Trichloroethane	0.868	0.826	4.8	90	0.00
33 S	1,2-Dichloroethane-d4	0.504	0.513	-1.8	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	91	0.00
35 S	Dibromofluoromethane	0.311	0.328	-5.5	92	0.00
36 T	1,1-Dichloropropene	0.447	0.445	0.4	89	0.00
37 T	Ethyl Acetate	0.141	0.126	10.6	81	0.00
38 T	Carbon Tetrachloride	0.446	0.456	-2.2	94	0.00
39 T	Methylcyclohexane	0.551	0.565	-2.5	89	0.00
40 TM	Benzene	1.303	1.251	4.0	87	0.00
41 T	Methacrylonitrile	0.082	0.071	13.4	72	0.00
42 TM	1,2-Dichloroethane	0.341	0.331	2.9	89	0.00
43 T	Isopropyl Acetate	0.270	0.260	3.7	86	0.00
44 TM	Trichloroethene	0.362	0.358	1.1	91	0.00
45 C	1,2-Dichloropropane	0.312	0.296	5.1#	87	0.00
46 T	Dibromomethane	0.177	0.163	7.9	83	0.00
47 T	Bromodichloromethane	0.455	0.436	4.2	88	0.00
48 T	Methyl methacrylate	0.137	0.132	3.6	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
 Data File : VD068730.D
 Acq On : 31 Mar 2021 11:01
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 01 00:41:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	89	0.00
50 S	Toluene-d8	1.199	1.302	-8.6	92	0.00
51 T	4-Methyl-2-Pentanone	0.141	0.138	2.1	88	0.00
52 CM	Toluene	0.825	0.808	2.1#	87	0.00
53 T	t-1,3-Dichloropropene	0.418	0.401	4.1	87	0.00
54 T	cis-1,3-Dichloropropene	0.497	0.468	5.8	86	0.00
55 T	1,1,2-Trichloroethane	0.249	0.231	7.2	87	0.00
56 T	Ethyl methacrylate	0.269	0.259	3.7	83	0.00
57 T	1,3-Dichloropropane	0.407	0.380	6.6	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.137	0.158	-15.3	98	0.00
59 T	2-Hexanone	0.094	0.099	-5.3	92	0.00
60 T	Dibromochloromethane	0.309	0.290	6.1	86	0.00
61 T	1,2-Dibromoethane	0.241	0.221	8.3	84	0.00
62 S	4-Bromofluorobenzene	0.389	0.414	-6.4	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.324	0.320	1.2	89	0.00
65 PM	Chlorobenzene	0.984	0.902	8.3	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.310	12.4	79	0.00
67 C	Ethyl Benzene	1.717	1.593	7.2#	82	0.00
68 T	m/p-Xylenes	0.669	0.675	-0.9	90	0.00
69 T	o-Xylene	0.620	0.604	2.6	86	0.00
70 T	Styrene	1.065	1.052	1.2	86	0.00
71 P	Bromoform	0.187	0.184	1.6	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.435	3.413	0.6	90	0.00
74 T	N-amyl acetate	0.573	0.528	7.9	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.607	0.577	4.9	89	0.00
76 T	1,2,3-Trichloropropane	0.416	0.362	13.0	85	0.00
77 T	Bromobenzene	0.806	0.755	6.3	86	0.00
78 T	n-propylbenzene	4.090	4.079	0.3	90	0.00
79 T	2-Chlorotoluene	2.232	2.212	0.9	91	0.00
80 T	1,3,5-Trimethylbenzene	2.778	2.685	3.3	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.183	0.177	3.3	86	0.00
82 T	4-Chlorotoluene	2.385	2.210	7.3	84	0.00
83 T	tert-Butylbenzene	2.368	2.389	-0.9	90	0.00
84 T	1,2,4-Trimethylbenzene	2.870	2.834	1.3	87	0.00
85 T	sec-Butylbenzene	3.442	3.493	-1.5	90	0.00
86 T	p-Isopropyltoluene	3.132	3.157	-0.8	89	0.00
87 T	1,3-Dichlorobenzene	1.607	1.501	6.6	86	0.00
88 T	1,4-Dichlorobenzene	1.580	1.496	5.3	87	0.00
89 T	n-Butylbenzene	2.974	2.986	-0.4	89	0.00
90 T	Hexachloroethane	0.564	0.593	-5.1	111	0.00
91 T	1,2-Dichlorobenzene	1.365	1.309	4.1	89	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.087	0.085	2.3	88	0.00
93 T	1,2,4-Trichlorobenzene	0.887	0.873	1.6	91	0.00
94 T	Hexachlorobutadiene	0.506	0.518	-2.4	92	0.00
95 T	Naphthalene	1.553	1.521	2.1	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD033121\
Data File : VD068730.D
Acq On : 31 Mar 2021 11:01
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 01 00:41:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.755	0.746	1.2	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6