

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040119\
 Data File : VD061472.D
 Acq On : 1 Apr 2019 16:54
 Operator : VA/AP
 Sample : K2178-07
 Misc : 5.78g/5ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC-16

Quant Time: Apr 02 02:00:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	520060	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.22	114	905107	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.36	117	569374	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.51	152	147869	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	241846	54.44	ug/l	0.02
Spiked Amount 50.000			Recovery	=	108.88%	
35) Dibromofluoromethane	6.93	113	315675	45.64	ug/l	0.03
Spiked Amount 50.000			Recovery	=	91.28%	
50) Toluene-d8	10.41	98	683789	41.22	ug/l	0.02
Spiked Amount 50.000			Recovery	=	82.44%	
62) 4-Bromofluorobenzene	13.56	95	216340	34.01	ug/l	0.01
Spiked Amount 50.000			Recovery	=	68.02%	
Target Compounds						
16) Acetone	3.23	43	74859	65.549	ug/l	97
20) Methylene Chloride	3.83	84	56362	7.165	ug/l	97
52) Toluene	10.50	92	24603	2.034	ug/l	98
68) m/p-Xylenes	12.66	106	9240	1.373	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040119\
Data File : VD061472.D
Acq On : 1 Apr 2019 16:54
Operator : VA/AP
Sample : K2178-07
Misc : 5.78g/5ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC-16

Quant Time: Apr 02 02:00:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

