

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040119\
 Data File : VD061475.D
 Acq On : 1 Apr 2019 18:19
 Operator : VA/AP
 Sample : K1689-03
 Misc : 6.32g/5ml/MSVOA D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-08-032919-A

Quant Time: Apr 02 02:07:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.04	168	578789	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.21	114	963647	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.37	117	728824	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.51	152	284309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	235218	47.57	ug/l	0.02
Spiked Amount 50.000			Recovery	=	95.14%	
35) Dibromofluoromethane	6.92	113	375627	51.01	ug/l	0.02
Spiked Amount 50.000			Recovery	=	102.02%	
50) Toluene-d8	10.41	98	737020	41.73	ug/l	0.02
Spiked Amount 50.000			Recovery	=	83.46%	
62) 4-Bromofluorobenzene	13.55	95	280245	41.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.76%	
Target Compounds						
16) Acetone	3.22	43	41777	32.869	ug/l	95
20) Methylene Chloride	3.82	84	40863	3.972	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040119\
Data File : VD061475.D
Acq On : 1 Apr 2019 18:19
Operator : VA/AP
Sample : K1689-03
Misc : 6.32g/5ml/MSVOA D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-08-032919-A

Quant Time: Apr 02 02:07:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

