

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040119\
 Data File : VD061502.D
 Acq On : 2 Apr 2019 8:34
 Operator : VA/AP
 Sample : K2180-26
 Misc : 5.14g/5ml/MSVOA D/SOIL
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1-2

Quant Time: Apr 03 01:14:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	616990	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.20	114	1015278	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.35	117	787758	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	297114	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	263738	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
35) Dibromofluoromethane	6.92	113	406639	52.41	ug/l	0.02
Spiked Amount 50.000			Recovery	=	104.82%	
50) Toluene-d8	10.39	98	899387	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	13.54	95	324289	45.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.88%	
Target Compounds						
16) Acetone	3.21	43	14339	10.583	ug/l #	81
20) Methylene Chloride	3.81	84	42038	3.763	ug/l	96
40) Benzene	7.45	78	29380	1.321	ug/l	97
52) Toluene	10.49	92	214778	15.830	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040119\
Data File : VD061502.D
Acq On : 2 Apr 2019 8:34
Operator : VA/AP
Sample : K2180-26
Misc : 5.14g/5ml/MSVOA D/SOIL
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1-2

Quant Time: Apr 03 01:14:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

