

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040121\
 Data File : VD068756.D
 Acq On : 01 Apr 2021 14:07
 Operator : VA/SY
 Sample : M1886-05
 Misc : 6.66G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124018-A

Quant Time: Apr 02 01:03:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	31170	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	52721	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	59278	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	27911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	20460	65.10	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	130.20%
35) Dibromofluoromethane	7.920	113	17464	53.24	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.48%
50) Toluene-d8	10.338	98	63181	49.96	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.92%
62) 4-Bromofluorobenzene	12.632	95	23180	56.53	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.06%
Target Compounds						
52) Toluene	10.408	92	1410	1.62	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040121\
Data File : VD068756.D
Acq On : 01 Apr 2021 14:07
Operator : VA/SY
Sample : M1886-05
Misc : 6.66G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124018-A

Quant Time: Apr 02 01:03:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

