

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040121\
 Data File : VD068763.D
 Acq On : 01 Apr 2021 17:25
 Operator : VA/SY
 Sample : M1889-01
 Misc : 11.28G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-5-(0-4)

Quant Time: Apr 02 01:05:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	16574	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	26100	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	29603	50.00	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	14285	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	11282	67.51	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	135.02%
35) Dibromofluoromethane	7.914	113	7180	44.21	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.42%
50) Toluene-d8	10.338	98	32698	52.23	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.46%
62) 4-Bromofluorobenzene	12.632	95	12323	60.71	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	121.42%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040121\
Data File : VD068763.D
Acq On : 01 Apr 2021 17:25
Operator : VA/SY
Sample : M1889-01
Misc : 11.28G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-5-(0-4)

Quant Time: Apr 02 01:05:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

