

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040121\
 Data File : VD068765.D
 Acq On : 01 Apr 2021 18:22
 Operator : VA/SY
 Sample : M1887-01RE
 Misc : 5.19G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-03-033121RE

Quant Time: Apr 02 01:37:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 20 02:00:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	37410	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	64231	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	71791	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	35415	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	26833	71.13	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.26%
35) Dibromofluoromethane	7.914	113	22120	55.35	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.70%
50) Toluene-d8	10.343	98	79414	51.54	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.08%
62) 4-Bromofluorobenzene	12.631	95	31269	62.59	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.18%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040121\
Data File : VD068765.D
Acq On : 01 Apr 2021 18:22
Operator : VA/SY
Sample : M1887-01RE
Misc : 5.19G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-03-033121RE

Quant Time: Apr 02 01:37:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D031921S.M
Quant Title : SW846 8260
QLast Update : Sat Mar 20 02:00:55 2021
Response via : Initial Calibration

