

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040122\
 Data File : VD072508.D
 Acq On : 01 Apr 2022 17:47
 Operator : VA/SY
 Sample : N2217-09
 Misc : 5.45G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-8

Quant Time: Apr 02 02:39:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	419505	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	772806	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	738987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	339502	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	231732	45.664	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.320%
35) Dibromofluoromethane	7.908	113	248764	41.421	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.840%
50) Toluene-d8	10.332	98	941298	45.089	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.180%
62) 4-Bromofluorobenzene	12.626	95	320786	39.345	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.680%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040122\
Data File : VD072508.D
Acq On : 01 Apr 2022 17:47
Operator : VA/SY
Sample : N2217-09
Misc : 5.45G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-8

Quant Time: Apr 02 02:39:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

