

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040122\
 Data File : VD072509.D
 Acq On : 01 Apr 2022 18:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 02:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	80	0.00
2 T	Dichlorodifluoromethane	0.397	0.371	6.5	66	0.00
3 P	Chloromethane	0.367	0.377	-2.7	69	0.00
4 C	Vinyl Chloride	0.385	0.379	1.6#	65	0.00
5 T	Bromomethane	0.283	0.294	-3.9	73	0.00
6 T	Chloroethane	0.270	0.280	-3.7	71	0.00
7 T	Trichlorofluoromethane	0.861	0.852	1.0	68	0.00
8 T	Diethyl Ether	0.256	0.268	-4.7	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.574	0.548	4.5	69	0.00
10 T	Methyl Iodide	0.457	0.458	-0.2	64	0.00
11 T	Tert butyl alcohol	0.032	0.031	3.1	75	0.00
12 CM	1,1-Dichloroethene	0.449	0.460	-2.4#	69	0.00
13 T	Acrolein	0.014	0.011	21.4	98	0.00
14 T	Allyl chloride	0.721	0.710	1.5	67	0.00
15 T	Acrylonitrile	0.128	0.129	-0.8	76	-0.01
16 T	Acetone	0.113	0.091	19.5	60	0.00
17 T	Carbon Disulfide	0.795	0.921	-15.8	66	0.00
18 T	Methyl Acetate	0.318	0.287	9.7	78	0.00
19 T	Methyl tert-butyl Ether	1.311	1.363	-4.0	76	0.00
20 T	Methylene Chloride	0.671	0.631	6.0	75	0.00
21 T	trans-1,2-Dichloroethene	0.502	0.553	-10.2	72	0.00
22 T	Diisopropyl ether	1.804	1.843	-2.2	73	0.00
23 T	Vinyl Acetate	0.871	0.967	-11.0	73	0.00
24 P	1,1-Dichloroethane	1.087	1.103	-1.5	74	-0.01
25 T	2-Butanone	0.156	0.147	5.8	69	0.00
26 T	2,2-Dichloropropane	0.966	0.874	9.5	66	0.00
27 T	cis-1,2-Dichloroethene	0.682	0.707	-3.7	73	0.00
28 T	Bromochloromethane	0.466	0.469	-0.6	77	0.00
29 T	Tetrahydrofuran	0.094	0.099	-5.3	75	0.00
30 C	Chloroform	1.223	1.267	-3.6#	77	0.00
31 T	Cyclohexane	0.791	0.762	3.7	66	0.00
32 T	1,1,1-Trichloroethane	1.037	1.064	-2.6	74	0.00
33 S	1,2-Dichloroethane-d4	0.605	0.586	3.1	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	80	0.00
35 S	Dibromofluoromethane	0.389	0.374	3.9	76	0.00
36 T	1,1-Dichloropropene	0.455	0.477	-4.8	70	0.00
37 T	Ethyl Acetate	0.206	0.216	-4.9	77	0.00
38 T	Carbon Tetrachloride	0.535	0.551	-3.0	72	0.00
39 T	Methylcyclohexane	0.498	0.509	-2.2	66	0.00
40 TM	Benzene	1.357	1.450	-6.9	73	0.00
41 T	Methacrylonitrile	0.112	0.113	-0.9	80	0.00
42 TM	1,2-Dichloroethane	0.415	0.453	-9.2	78	0.00
43 T	Isopropyl Acetate	0.395	0.425	-7.6	77	0.00
44 TM	Trichloroethene	0.385	0.406	-5.5	73	0.00
45 C	1,2-Dichloropropane	0.377	0.393	-4.2#	75	0.00
46 T	Dibromomethane	0.198	0.221	-11.6	79	0.00
47 T	Bromodichloromethane	0.569	0.607	-6.7	80	0.00
48 T	Methyl methacrylate	0.192	0.189	1.6	73	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040122\
 Data File : VD072509.D
 Acq On : 01 Apr 2022 18:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 02 02:39:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	75	0.00
50 S	Toluene-d8	1.351	1.275	5.6	75	0.00
51 T	4-Methyl-2-Pentanone	0.212	0.228	-7.5	78	0.00
52 CM	Toluene	0.902	0.981	-8.8#	74	0.00
53 T	t-1,3-Dichloropropene	0.496	0.531	-7.1	76	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.604	-3.1	73	0.00
55 T	1,1,2-Trichloroethane	0.300	0.320	-6.7	79	0.00
56 T	Ethyl methacrylate	0.339	0.383	-13.0	77	0.00
57 T	1,3-Dichloropropane	0.496	0.532	-7.3	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.148	0.150	-1.4	74	0.00
59 T	2-Hexanone	0.144	0.153	-6.3	75	0.00
60 T	Dibromochloromethane	0.387	0.425	-9.8	82	0.00
61 T	1,2-Dibromoethane	0.267	0.303	-13.5	83	0.00
62 S	4-Bromofluorobenzene	0.528	0.503	4.7	78	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.327	0.325	0.6	73	0.00
65 PM	Chlorobenzene	1.118	1.142	-2.1	77	0.00
66 T	1,1,1,2-Tetrachloroethane	0.447	0.462	-3.4	79	0.00
67 C	Ethyl Benzene	1.951	2.004	-2.7#	74	0.00
68 T	m/p-Xylenes	0.740	0.781	-5.5	76	0.00
69 T	o-Xylene	0.714	0.753	-5.5	77	0.00
70 T	Styrene	1.258	1.342	-6.7	78	0.00
71 P	Bromoform	0.237	0.252	-6.3	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	4.038	3.864	4.3	74	0.00
74 T	N-amyl acetate	0.847	0.825	2.6	77	0.00
75 P	1,1,2,2-Tetrachloroethane	0.765	0.751	1.8	84	0.00
76 T	1,2,3-Trichloropropane	0.545	0.555	-1.8	76	0.00
77 T	Bromobenzene	0.905	0.912	-0.8	80	0.00
78 T	n-propylbenzene	4.871	4.727	3.0	75	0.00
79 T	2-Chlorotoluene	2.849	2.776	2.6	77	0.00
80 T	1,3,5-Trimethylbenzene	3.458	3.368	2.6	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.222	0.204	8.1	74	0.00
82 T	4-Chlorotoluene	2.967	2.882	2.9	78	0.00
83 T	tert-Butylbenzene	2.994	2.868	4.2	76	0.00
84 T	1,2,4-Trimethylbenzene	3.382	3.319	1.9	78	0.00
85 T	sec-Butylbenzene	4.510	4.237	6.1	73	0.00
86 T	p-Isopropyltoluene	3.686	3.552	3.6	76	0.00
87 T	1,3-Dichlorobenzene	1.875	1.832	2.3	80	0.00
88 T	1,4-Dichlorobenzene	1.890	1.809	4.3	80	0.00
89 T	n-Butylbenzene	3.512	3.346	4.7	75	0.00
90 T	Hexachloroethane	0.755	0.719	4.8	79	0.00
91 T	1,2-Dichlorobenzene	1.654	1.639	0.9	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.113	0.117	-3.5	87	0.00
93 T	1,2,4-Trichlorobenzene	1.025	0.985	3.9	77	0.00
94 T	Hexachlorobutadiene	0.574	0.530	7.7	77	0.00
95 T	Naphthalene	1.992	1.923	3.5	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040122\
Data File : VD072509.D
Acq On : 01 Apr 2022 18:15
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Apr 02 02:39:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.895	0.871	2.7	79	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6