

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040219\
 Data File : VD061518.D
 Acq On : 2 Apr 2019 17:02
 Operator : VA/AP
 Sample : K2183-02RE
 Misc : 4.56g/5ml/MSVOA D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-0D015-03-002RE

Manual Integrations
 APPROVED

MMDadoda
 4/3/2019 1:02:57 PM

Quant Time: Apr 03 08:12:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.02	168	498500	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.21	114	889398	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.36	117	715896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	338479	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.44	65	205774	48.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.64%	
35) Dibromofluoromethane	6.92	113	1194	0.18	ug/l	0.02
Spiked Amount 50.000			Recovery	=	0.36%	
50) Toluene-d8	10.40	98	697264	42.77	ug/l	0.02
Spiked Amount 50.000			Recovery	=	85.54%	
62) 4-Bromofluorobenzene	13.55	95	294524	47.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.24%	
Target Compounds						
16) Acetone	3.28	43	145425m	132.846	ug/l	Qvalue
20) Methylene Chloride	3.79	84	54678	7.276	ug/l	97
25) 2-Butanone	6.10	43	17192m	12.485	ug/l	
40) Benzene	7.44	78	36333	1.865	ug/l	92
52) Toluene	10.49	92	40913	3.442	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040219\
Data File : VD061518.D
Acq On : 2 Apr 2019 17:02
Operator : VA/AP
Sample : K2183-02RE
Misc : 4.56g/5ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-0D015-03-002RE

Manual Integrations
APPROVED

MMDadoda
4/3/2019 1:02:57 PM

Quant Time: Apr 03 08:12:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

