

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040219\
 Data File : VD061525.D
 Acq On : 2 Apr 2019 20:16
 Operator : VA/AP
 Sample : K2178-07RE
 Misc : 4.88g/5ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC-16RE

Quant Time: Apr 03 08:47:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	401481	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.22	114	743333	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.38	117	481348	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	14.53	152	107465	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	217736	63.49	ug/l	0.02
Spiked Amount 50.000			Recovery	=	126.98%	
35) Dibromofluoromethane	6.93	113	32804	5.77	ug/l	0.03
Spiked Amount 50.000			Recovery	=	11.54%	
50) Toluene-d8	10.42	98	578012	42.43	ug/l	0.04
Spiked Amount 50.000			Recovery	=	84.86%	
62) 4-Bromofluorobenzene	13.56	95	178528	34.17	ug/l	0.02
Spiked Amount 50.000			Recovery	=	68.34%	
Target Compounds						
16) Acetone	3.21	43	113346	128.563	ug/l	96
20) Methylene Chloride	3.82	84	31617	4.661	ug/l	91
52) Toluene	10.51	92	23800	2.396	ug/l	98
68) m/p-Xylenes	12.68	106	7355	1.293	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040219\
Data File : VD061525.D
Acq On : 2 Apr 2019 20:16
Operator : VA/AP
Sample : K2178-07RE
Misc : 4.88g/5ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC-16RE

Quant Time: Apr 03 08:47:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

