

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040219\
 Data File : VD061527.D
 Acq On : 2 Apr 2019 21:12
 Operator : VA/AP
 Sample : K2182-19RE
 Misc : 4.99g/5ml/MSVOA D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S7C(16)RE

Quant Time: Apr 03 08:56:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	98389	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.22	114	183261	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.36	117	148070	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	55060	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	84098	100.06	ug/l	0.03
Spiked Amount 50.000			Recovery	=	200.12%	
35) Dibromofluoromethane	6.93	113	90317	64.49	ug/l	0.03
Spiked Amount 50.000			Recovery	=	128.98%	
50) Toluene-d8	10.41	98	157645	46.93	ug/l	0.03
Spiked Amount 50.000			Recovery	=	93.86%	
62) 4-Bromofluorobenzene	13.55	95	66153	51.36	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.72%	
Target Compounds						
16) Acetone	3.23	43	73626	340.768	ug/l	99
20) Methylene Chloride	3.84	84	15890	11.656	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040219\
Data File : VD061527.D
Acq On : 2 Apr 2019 21:12
Operator : VA/AP
Sample : K2182-19RE
Misc : 4.99g/5ml/MSVOA D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S7C(16)RE

Quant Time: Apr 03 08:56:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

