

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040219\
 Data File : VD061529.D
 Acq On : 2 Apr 2019 22:07
 Operator : VA/AP
 Sample : K2186-23
 Misc : 7.08g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-TRACK-78-79

Quant Time: Apr 03 09:00:10 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	559998	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.23	114	892943	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.37	117	757698	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	332930	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	165369	34.57	ug/l	0.03
Spiked Amount 50.000			Recovery	=	69.14%	
35) Dibromofluoromethane	6.94	113	279437	40.95	ug/l	0.04
Spiked Amount 50.000			Recovery	=	81.90%	
50) Toluene-d8	10.41	98	621273	37.96	ug/l	0.03
Spiked Amount 50.000			Recovery	=	75.92%	
62) 4-Bromofluorobenzene	13.55	95	250294	39.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.76%	
Target Compounds						
16) Acetone	3.23	43	61668	50.147	ug/l	93
20) Methylene Chloride	3.84	84	25605	1.869	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040219\
Data File : VD061529.D
Acq On : 2 Apr 2019 22:07
Operator : VA/AP
Sample : K2186-23
Misc : 7.08g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-TRACK-78-79

Quant Time: Apr 03 09:00:10 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

