

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040219\
 Data File : VD061544.D
 Acq On : 3 Apr 2019 7:47
 Operator : VA/AP
 Sample : K1695-08
 Misc : 5.63g/5ml/MSVOA D/SOIL
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DPC-91723-SO-TP5-B

Quant Time: Apr 03 09:27:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	24135	50.00	ug/l	0.05
34) 1,4-Difluorobenzene	8.23	114	47482	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.38	117	47149	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	23799	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	26709	129.55	ug/l	0.04
Spiked Amount 50.000			Recovery	=	259.10%	
35) Dibromofluoromethane	6.94	113	28002	77.17	ug/l	0.04
Spiked Amount 50.000			Recovery	=	154.34%	
50) Toluene-d8	10.42	98	51864	59.59	ug/l	0.03
Spiked Amount 50.000			Recovery	=	119.18%	
62) 4-Bromofluorobenzene	13.56	95	27119	81.26	ug/l	0.01
Spiked Amount 50.000			Recovery	=	162.52%	
Target Compounds						
16) Acetone	3.24	43	12692	239.474	ug/l #	86
20) Methylene Chloride	3.85	84	3340	9.702	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040219\
Data File : VD061544.D
Acq On : 3 Apr 2019 7:47
Operator : VA/AP
Sample : K1695-08
Misc : 5.63g/5ml/MSVOA D/SOIL
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DPC-91723-SO-TP5-B

Quant Time: Apr 03 09:27:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

