

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040219\
 Data File : VD061548.D
 Acq On : 3 Apr 2019 9:38
 Operator : VA/AP
 Sample : K2219-01
 Misc : 3.82g/5ml/MSVOA D/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-1-2-COMP

Quant Time: Apr 04 01:10:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	468339	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.23	114	781766	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.37	117	631714	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	216348	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	222695	55.66	ug/l	0.03
Spiked Amount 50.000			Recovery	=	111.32%	
35) Dibromofluoromethane	6.94	113	313761	52.52	ug/l	0.04
Spiked Amount 50.000			Recovery	=	105.04%	
50) Toluene-d8	10.41	98	670692	46.81	ug/l	0.03
Spiked Amount 50.000			Recovery	=	93.62%	
62) 4-Bromofluorobenzene	13.56	95	222608	40.51	ug/l	0.01
Spiked Amount 50.000			Recovery	=	81.02%	
Target Compounds						
16) Acetone	3.24	43	30565	29.719	ug/l	Qvalue 100
20) Methylene Chloride	3.84	84	83775	13.125	ug/l	97
52) Toluene	10.51	92	14581	1.396	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040219\
Data File : VD061548.D
Acq On : 3 Apr 2019 9:38
Operator : VA/AP
Sample : K2219-01
Misc : 3.82g/5ml/MSVOA D/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-1-2-COMP

Quant Time: Apr 04 01:10:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

