

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040219\
 Data File : VD061549.D
 Acq On : 3 Apr 2019 10:09
 Operator : VA/AP
 Sample : K2219-02
 Misc : 5.66g/5ml/MSVOA D/SOIL
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2-3-COMP

Quant Time: Apr 04 01:12:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	409204	50.00	ug/l	0.05
34) 1,4-Difluorobenzene	8.24	114	712801	50.00	ug/l	0.05
63) Chlorobenzene-d5	12.40	117	561081	50.00	ug/l	0.04
72) 1,4-Dichlorobenzene-d4	14.54	152	210362	50.00	ug/l	0.03
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	238283	68.17	ug/l	0.04
Spiked Amount 50.000			Recovery	=	136.34%	
35) Dibromofluoromethane	6.95	113	316125	58.04	ug/l	0.05
Spiked Amount 50.000			Recovery	=	116.08%	
50) Toluene-d8	10.44	98	597026	45.70	ug/l	0.05
Spiked Amount 50.000			Recovery	=	91.40%	
62) 4-Bromofluorobenzene	13.58	95	234984	46.90	ug/l	0.03
Spiked Amount 50.000			Recovery	=	93.80%	
Target Compounds						
20) Methylene Chloride	3.83	84	63787	11.181	ug/l	91
40) Benzene	7.49	78	25502	1.633	ug/l	100
52) Toluene	10.54	92	105997	11.128	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040219\
Data File : VD061549.D
Acq On : 3 Apr 2019 10:09
Operator : VA/AP
Sample : K2219-02
Misc : 5.66g/5ml/MSVOA D/SOIL
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2-3-COMP

Quant Time: Apr 04 01:12:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

