

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040319\
 Data File : VD061556.D
 Acq On : 3 Apr 2019 15:17
 Operator : VA/AP
 Sample : VD0403SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0403

Quant Time: Apr 04 02:38:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	639720	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	8.22	114	1025345	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.37	117	841664	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	384158	50.00	ug/l	0.01

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.46	65	281447	51.50	ug/l	0.02
Spiked Amount 50.000			Recovery	=	103.00%	
35) Dibromofluoromethane	6.93	113	429580	54.82	ug/l	0.03
Spiked Amount 50.000			Recovery	=	109.64%	
50) Toluene-d8	10.42	98	1024667	54.52	ug/l	0.03
Spiked Amount 50.000			Recovery	=	109.04%	
62) 4-Bromofluorobenzene	13.56	95	428196	59.42	ug/l	0.01
Spiked Amount 50.000			Recovery	=	118.84%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040319\
Data File : VD061556.D
Acq On : 3 Apr 2019 15:17
Operator : VA/AP
Sample : VD0403SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0403

Quant Time: Apr 04 02:38:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

