

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040419\
 Data File : VD061582.D
 Acq On : 4 Apr 2019 16:44
 Operator : VA/AP
 Sample : K2199-03RE
 Misc : 5.05g/5ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-040219-BRE

Quant Time: Apr 05 01:29:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 28 09:22:37 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.03	168	499176	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.20	114	729445	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.35	117	168147	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.51	152	190074	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.43	65	175741	41.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.42%	
35) Dibromofluoromethane	6.91	113	278991	50.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.10%	
50) Toluene-d8	10.39	98	80325	6.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	12.02%	
62) 4-Bromofluorobenzene	13.54	95	180821	35.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	70.54%	
Target Compounds						
16) Acetone	3.20	43	11125	10.149	ug/l #	76
20) Methylene Chloride	3.81	84	32902	3.575	ug/l	91
69) o-Xylene	13.04	106	2946	1.528	ug/l	90
95) Naphthalene	16.11	128	5164	1.365	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040419\
Data File : VD061582.D
Acq On : 4 Apr 2019 16:44
Operator : VA/AP
Sample : K2199-03RE
Misc : 5.05g/5ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-040219-BRE

Quant Time: Apr 05 01:29:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D032719S.M
Quant Title : SW846 8260
QLast Update : Thu Mar 28 09:22:37 2019
Response via : Initial Calibration

