

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040422\
 Data File : VD072512.D
 Acq On : 04 Apr 2022 11:20
 Operator : VA/SY
 Sample : VD0404SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0402

Quant Time: Apr 05 03:18:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	540055	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	988753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	926133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	430683	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	302094	46.241	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.480%
35) Dibromofluoromethane	7.914	113	325079	42.306	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.620%
50) Toluene-d8	10.332	98	1185285	44.376	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.760%
62) 4-Bromofluorobenzene	12.626	95	409268	39.234	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040422\
Data File : VD072512.D
Acq On : 04 Apr 2022 11:20
Operator : VA/SY
Sample : VD0404SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0402

Quant Time: Apr 05 03:18:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

