

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040522\
 Data File : VD072533.D
 Acq On : 05 Apr 2022 14:20
 Operator : VA/SY
 Sample : N2256-03
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-03-040422

Quant Time: Apr 06 01:12:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	423874	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	797844	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	738596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	328511	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	242451	47.284	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.560%
35) Dibromofluoromethane	7.908	113	262404	42.321	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.640%
50) Toluene-d8	10.332	98	964181	44.736	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.480%
62) 4-Bromofluorobenzene	12.620	95	320263	38.048	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.100%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	4.414	76	2497	3.907	ug/l #	83
20) Methylene Chloride	4.961	84	118178	20.762	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040522\
Data File : VD072533.D
Acq On : 05 Apr 2022 14:20
Operator : VA/SY
Sample : N2256-03
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-03-040422

Quant Time: Apr 06 01:12:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

