

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040522\
 Data File : VD072538.D
 Acq On : 05 Apr 2022 16:42
 Operator : VA/SY
 Sample : N2254-12
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-14A-1A-4-GR-5

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/06/2022
 Supervised By : Mahesh Dadoda 04/06/2022

Quant Time: Apr 06 01:14:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	703968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1237092	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1117770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	478754	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	351957	41.330	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	82.660%		
35) Dibromofluoromethane	7.908	113	385596	40.108	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	80.220%		
50) Toluene-d8	10.332	98	1584673	47.419	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	94.840%		
62) 4-Bromofluorobenzene	12.620	95	516565	39.579	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	79.160%		
Target Compounds					Qvalue	
17) Carbon Disulfide	4.403	76	3324m	3.851	ug/l	
31) Cyclohexane	7.979	56	133632	12.001	ug/l #	78
39) Methylcyclohexane	9.350	83	1707242	115.332	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040522\
Data File : VD072538.D
Acq On : 05 Apr 2022 16:42
Operator : VA/SY
Sample : N2254-12
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-14A-1A-4-GR-5

Quant Time: Apr 06 01:14:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/06/2022
Supervised By : Mahesh Dadoda 04/06/2022

