

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040522\
 Data File : VD072540.D
 Acq On : 05 Apr 2022 17:39
 Operator : VA/SY
 Sample : N2252-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-01-040422

Quant Time: Apr 06 01:14:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	39363	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	69411	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	74186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	36594	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	36098	75.809	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	151.620%
35) Dibromofluoromethane	7.908	113	27744	51.433	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.860%
50) Toluene-d8	10.332	98	89264	47.606	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.220%
62) 4-Bromofluorobenzene	12.620	95	34527	47.149	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.300%
Target Compounds						
39) Methylcyclohexane	9.350	83	29846	37.915	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040522\
Data File : VD072540.D
Acq On : 05 Apr 2022 17:39
Operator : VA/SY
Sample : N2252-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-01-040422

Quant Time: Apr 06 01:14:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

