

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040523\
 Data File : VD075578.D
 Acq On : 05 Apr 2023 14:26
 Operator : KP/SY
 Sample : 02193-03
 Misc : 4.39G/10.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C11C6

Quant Time: Apr 05 23:38:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM033123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Apr 01 00:08:20 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	440565	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	290946	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	106866	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	228692	22.295	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.160%
7) Chloroethane-d5	2.805	69	199481	21.854	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.400%
11) 1,1-Dichloroethene-d2	3.922	63	164326	19.861	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	79.440%
21) 2-Butanone-d5	6.993	46	26852	40.198	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.400%
24) Chloroform-d	7.569	84	252031	24.264	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	97.040%
26) 1,2-Dichloroethane-d4	8.234	65	100652	21.866	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	87.480%
32) Benzene-d6	8.205	84	546225	32.113	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	128.440%
36) 1,2-Dichloropropane-d6	9.210	67	145081	32.005	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	128.040%#
41) Toluene-d8	10.269	98	398769	25.002	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.000%
43) trans-1,3-Dichloroprop...	10.528	79	43914	23.239	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.960%
47) 2-Hexanone-d5	10.881	63	19372	36.114	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	72.220%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	74342	21.892	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.560%
66) 1,2-Dichlorobenzene-d4	13.816	152	95189	25.967	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.880%
Target Compounds						Qvalue
13) Acetone	4.028	43	6699	9.562	ug/L	92
16) Methylene chloride	4.805	84	7057	1.165	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040523\
Data File : VD075578.D
Acq On : 05 Apr 2023 14:26
Operator : KP/SY
Sample : 02193-03
Misc : 4.39G/10.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C11C6

Quant Time: Apr 05 23:38:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM033123SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Apr 01 00:08:20 2023
Response via : Initial Calibration

