


Methyl Iodide

Response Ratio



$$\text{Response} = 6.367\text{e-}001 * \text{Amt} - 2.861\text{e-}002$$


Coef of Det (r^2) = 0.997752 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D040721S.M

Calibration Table Last Updated: Thu Apr 08 03:50:31 2021

Tert butyl alcohol


Response Ratio



Response = 2.470e-002 * Amt + 1.675e-002
 Coef of Det (r^2) = 0.999114 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D040721S.M
 Calibration Table Last Updated: Thu Apr 08 03:50:31 2021

Methylene Chloride

Response Ratio



$$\text{Response} = 5.357\text{e-}001 * \text{Amt} + 5.704\text{e-}002$$


Coef of Det (r^2) = 0.998238 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D040721S.M

Calibration Table Last Updated: Thu Apr 08 03:50:31 2021

2-Chloroethyl Vinyl ether

Response Ratio



$$\text{Response} = 1.638\text{e-}001 * \text{Amt} - 5.567\text{e-}002$$

Coef of Det (r^2) = 0.997555 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA D\Method\82D040721S.M

Calibration Table Last Updated: Thu Apr 08 03:50:31 2021