

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040819\
 Data File : VD061660.D
 Acq On : 8 Apr 2019 19:46
 Operator : VA/AP
 Sample : K2305-02
 Misc : 4.09g/5ml/MSVOA D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 FK-EF-02-001-ABCD

Quant Time: Apr 09 02:06:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	188136	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.23	114	309307	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.38	117	275236	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	129140	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	103709	72.72	ug/l	0.01
Spiked Amount 50.000			Recovery	=	145.44%	
35) Dibromofluoromethane	6.93	113	129560	60.46	ug/l	0.01
Spiked Amount 50.000			Recovery	=	120.92%	
50) Toluene-d8	10.42	98	257619	47.85	ug/l	0.01
Spiked Amount 50.000			Recovery	=	95.70%	
62) 4-Bromofluorobenzene	13.56	95	119348	68.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	136.16%	
Target Compounds						
16) Acetone	3.22	43	21756	51.035	ug/l	97
20) Methylene Chloride	3.82	84	37081	15.329	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040819\
Data File : VD061660.D
Acq On : 8 Apr 2019 19:46
Operator : VA/AP
Sample : K2305-02
Misc : 4.09g/5ml/MSVOA D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
FK-EF-02-001-ABCD

Quant Time: Apr 09 02:06:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

