

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040819\
 Data File : VD061661.D
 Acq On : 8 Apr 2019 20:16
 Operator : VA/AP
 Sample : K2294-02
 Misc : 5.56g/5ml/MSVOA D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-14-S2

Quant Time: Apr 09 02:11:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.06	168	330199	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	8.23	114	459552	50.00	ug/l	0.02
63) Chlorobenzene-d5	12.37	117	317983	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	14.52	152	75255	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	110328	44.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.16%	
35) Dibromofluoromethane	6.94	113	175443	55.11	ug/l	0.02
Spiked Amount 50.000			Recovery	=	110.22%	
50) Toluene-d8	10.42	98	328293	41.04	ug/l	0.02
Spiked Amount 50.000			Recovery	=	82.08%	
62) 4-Bromofluorobenzene	13.56	95	77739	24.75	ug/l	0.01
Spiked Amount 50.000			Recovery	=	49.50%	
Target Compounds						
16) Acetone	3.23	43	62318	83.291	ug/l	97
17) Carbon Disulfide	3.39	76	15928	1.424	ug/l	98
20) Methylene Chloride	3.83	84	25125	5.918	ug/l	95
40) Benzene	7.46	78	13952	1.438	ug/l #	88
95) Naphthalene	16.13	128	5836	3.507	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040819\
Data File : VD061661.D
Acq On : 8 Apr 2019 20:16
Operator : VA/AP
Sample : K2294-02
Misc : 5.56g/5ml/MSVOA D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-14-S2

Quant Time: Apr 09 02:11:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

