

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040819\
 Data File : VD061664.D
 Acq On : 8 Apr 2019 21:47
 Operator : VA/AP
 Sample : K2317-15
 Misc : 5.11g/5ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COMP-2

Quant Time: Apr 09 02:39:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	318288	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	8.22	114	521407	50.00	ug/l	0.01
63) Chlorobenzene-d5	12.38	117	364373	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	149110	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.47	65	154229	63.93	ug/l	0.01
Spiked Amount 50.000			Recovery	=	127.86%	
35) Dibromofluoromethane	6.93	113	219137	60.67	ug/l	0.01
Spiked Amount 50.000			Recovery	=	121.34%	
50) Toluene-d8	10.42	98	408057	44.96	ug/l	0.01
Spiked Amount 50.000			Recovery	=	89.92%	
62) 4-Bromofluorobenzene	13.56	95	165219	52.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.58%	
Target Compounds						
16) Acetone	3.21	43	8438	11.700	ug/l	93
20) Methylene Chloride	3.82	84	47203	11.534	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040819\
Data File : VD061664.D
Acq On : 8 Apr 2019 21:47
Operator : VA/AP
Sample : K2317-15
Misc : 5.11a/5ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COMP-2

Quant Time: Apr 09 02:39:53 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

