


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD040819\
 Data File : VD061646.D
 Acq On : 8 Apr 2019 12:27
 Operator : VA/AP
 Sample : BFB
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Title : SW846 8260
 Last Update : Mon Apr 08 10:04:03 2019



AutoFind: Scans 1232, 1233, 1234; Background Corrected with Scan 1228

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.2	15117	PASS
75	95	30	60	39.0	34386	PASS
95	95	100	100	100.0	88082	PASS
96	95	5	9	6.4	5673	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.2	75904	PASS
175	174	5	9	7.0	5320	PASS
176	174	95	101	96.6	73290	PASS
177	176	5	9	6.7	4895	PASS