

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
 Data File : VD068786.D
 Acq On : 08 Apr 2021 13:48
 Operator : VA/SY
 Sample : M1979-13
 Misc : 5.17G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-13

Quant Time: Apr 09 01:01:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	284710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	523319	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.643	117	552179	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	310735	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.338	65	131456	43.69	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	87.38%
35) Dibromofluoromethane	7.920	113	153028	44.96	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	89.92%
50) Toluene-d8	10.343	98	677483	52.76	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.52%
62) 4-Bromofluorobenzene	12.632	95	318539	74.69	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	149.38%#
Target Compounds						
					Qvalue	
16) Acetone	4.173	43	3810	9.14	ug/l #	73
17) Carbon Disulfide	4.414	76	9339	1.01	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
Data File : VD068786.D
Acq On : 08 Apr 2021 13:48
Operator : VA/SY
Sample : M1979-13
Misc : 5.17G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-13

Quant Time: Apr 09 01:01:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

