

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
 Data File : VD068787.D
 Acq On : 08 Apr 2021 14:16
 Operator : VA/SY
 Sample : M1979-14
 Misc : 5.52G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-14

Quant Time: Apr 09 01:02:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	427710	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	754631	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	760857	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	358900	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	183750	40.65	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.30%
35) Dibromofluoromethane	7.920	113	225511	45.94	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.88%
50) Toluene-d8	10.338	98	930253	50.24	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.48%
62) 4-Bromofluorobenzene	12.632	95	346225	56.30	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.60%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	7580	12.11	ug/l #	87
39) Methylcyclohexane	9.355	83	11190	1.26	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
Data File : VD068787.D
Acq On : 08 Apr 2021 14:16
Operator : VA/SY
Sample : M1979-14
Misc : 5.52G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-14

Quant Time: Apr 09 01:02:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

