

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
 Data File : VD068788.D
 Acq On : 08 Apr 2021 14:44
 Operator : VA/SY
 Sample : M1979-15
 Misc : 4.44G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-15

Quant Time: Apr 09 01:02:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	401003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	715452	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	734101	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	368358	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	183269	43.24	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.48%
35) Dibromofluoromethane	7.920	113	216189	46.46	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.92%
50) Toluene-d8	10.338	98	887521	50.56	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.12%
62) 4-Bromofluorobenzene	12.632	95	333811	57.25	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.50%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
Data File : VD068788.D
Acq On : 08 Apr 2021 14:44
Operator : VA/SY
Sample : M1979-15
Misc : 4.44G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-15

Quant Time: Apr 09 01:02:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

