

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
 Data File : VD068794.D
 Acq On : 08 Apr 2021 17:32
 Operator : VA/SY
 Sample : M1979-21
 Misc : 5.14G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-21

Quant Time: Apr 09 01:03:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.985	168	409945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	759696	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	823080	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	394810	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	196463	45.35	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.70%
35) Dibromofluoromethane	7.920	113	241188	48.81	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.62%
50) Toluene-d8	10.338	98	976531	52.39	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.78%
62) 4-Bromofluorobenzene	12.632	95	389945	62.99	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	125.98%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040821\
Data File : VD068794.D
Acq On : 08 Apr 2021 17:32
Operator : VA/SY
Sample : M1979-21
Misc : 5.14G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-21

Quant Time: Apr 09 01:03:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

