

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
 Data File : VD072560.D
 Acq On : 08 Apr 2022 15:31
 Operator : VA/SY
 Sample : N2317-05
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NORTH-1

Quant Time: Apr 09 00:36:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	497771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	916224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	854309	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	385964	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	285865	47.474	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	94.940%
35) Dibromofluoromethane	7.908	113	299177	42.017	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.040%
50) Toluene-d8	10.332	98	1110662	44.874	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.740%
62) 4-Bromofluorobenzene	12.620	95	378723	39.180	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
Data File : VD072560.D
Acq On : 08 Apr 2022 15:31
Operator : VA/SY
Sample : N2317-05
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NORTH-1

Quant Time: Apr 09 00:36:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

