

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
 Data File : VD072562.D
 Acq On : 08 Apr 2022 16:28
 Operator : VA/SY
 Sample : N2317-11
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NORTH-4

Quant Time: Apr 09 00:36:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	58721	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	103462	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	106065	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	48682	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	53453	75.249	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	150.500%
35) Dibromofluoromethane	7.914	113	45470	56.551	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.100%
50) Toluene-d8	10.332	98	124345	44.490	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.980%
62) 4-Bromofluorobenzene	12.620	95	44642	40.898	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	81.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
Data File : VD072562.D
Acq On : 08 Apr 2022 16:28
Operator : VA/SY
Sample : N2317-11
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NORTH-4

Quant Time: Apr 09 00:36:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

