

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
 Data File : VD072563.D
 Acq On : 08 Apr 2022 16:56
 Operator : VA/SY
 Sample : N2317-13
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NORTH-5

Quant Time: Apr 09 00:36:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	460894	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	844035	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	782112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	338194	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	265144	47.556	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.120%
35) Dibromofluoromethane	7.908	113	281309	42.887	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.780%
50) Toluene-d8	10.332	98	1021395	44.797	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.600%
62) 4-Bromofluorobenzene	12.626	95	333960	37.504	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	75.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
Data File : VD072563.D
Acq On : 08 Apr 2022 16:56
Operator : VA/SY
Sample : N2317-13
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NORTH-5

Quant Time: Apr 09 00:36:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

