

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
 Data File : VD072565.D
 Acq On : 08 Apr 2022 17:53
 Operator : VA/SY
 Sample : N2317-17
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP1-A

Quant Time: Apr 09 00:37:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 02:44:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	436688	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	824101	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	773232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	342254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	261803	49.560	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.120%
35) Dibromofluoromethane	7.908	113	271790	42.438	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.880%
50) Toluene-d8	10.332	98	992268	44.572	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	89.140%
62) 4-Bromofluorobenzene	12.620	95	333875	38.401	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	76.800%
Target Compounds						
16) Acetone	4.155	43	27654	28.074	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040822\
Data File : VD072565.D
Acq On : 08 Apr 2022 17:53
Operator : VA/SY
Sample : N2317-17
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP1-A

Quant Time: Apr 09 00:37:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D032922S.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 02:44:36 2022
Response via : Initial Calibration

