

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040919\
 Data File : VD061688.D
 Acq On : 9 Apr 2019 21:25
 Operator : VA/AP
 Sample : K2316-05
 Misc : 5.02g/5ml/MSVOA D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :

Quant Time: Apr 10 00:57:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.08	168	444631	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.25	114	764376	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.39	117	470346	50.00	ug/l	0.03
72) 1,4-Dichlorobenzene-d4	14.53	152	114800	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.49	65	184633	54.78	ug/l	0.03
Spiked Amount 50.000			Recovery	=	109.56%	
35) Dibromofluoromethane	6.96	113	283408	53.52	ug/l	0.04
Spiked Amount 50.000			Recovery	=	107.04%	
50) Toluene-d8	10.44	98	568091	42.70	ug/l	0.03
Spiked Amount 50.000			Recovery	=	85.40%	
62) 4-Bromofluorobenzene	13.58	95	167297	33.32	ug/l	0.03
Spiked Amount 50.000			Recovery	=	66.64%	
Target Compounds						
					Qvalue	
16) Acetone	3.24	43	54755	54.348	ug/l	97
17) Carbon Disulfide	3.41	76	40476	2.687	ug/l	100
20) Methylene Chloride	3.85	84	19841	3.471	ug/l	94
25) 2-Butanone	6.10	43	18465	15.898	ug/l	98
36) 1,1-Dichloropropene	7.08	75	31752	4.746	ug/l #	48
37) Ethyl Acetate	6.10	43	18465	6.777	ug/l #	68
38) Carbon Tetrachloride	7.08	117	42820	5.462	ug/l #	16
40) Benzene	7.49	78	48622	3.013	ug/l	99
51) 4-Methyl-2-Pentanone	10.39	43	21521	9.202	ug/l #	78
56) Ethyl methacrylate	11.23	69	3654	1.019	ug/l #	1
65) Chlorobenzene	12.43	112	25815	2.768	ug/l #	87
67) Ethyl Benzene	12.55	91	23225	1.635	ug/l	92
68) m/p-Xylenes	12.68	106	14274	2.607	ug/l	93
69) o-Xylene	13.07	106	11339	2.145	ug/l	97
73) Isopropylbenzene	13.42	105	101303	13.244	ug/l	100
74) N-amyl acetate	13.24	43	15827	7.615	ug/l #	54
76) 1,2,3-Trichloropropane	13.58	75	62292	39.883	ug/l #	47
78) n-propylbenzene	13.78	91	25332	2.769	ug/l #	54
79) 2-Chlorotoluene	13.86	91	76247	14.572	ug/l	92
80) 1,3,5-Trimethylbenzene	13.95	105	10910	1.797	ug/l	93
81) trans-1,4-Dichloro-2-buten	13.58	75	62483	131.834	ug/l #	20
83) tert-Butylbenzene	14.21	119	26351	3.600	ug/l	96
84) 1,2,4-Trimethylbenzene	14.26	105	20999	3.427	ug/l	96
92) 1,2-Dibromo-3-Chloropropan	15.35	75	822	4.307	ug/l #	14
95) Naphthalene	16.14	128	17107	6.738	ug/l #	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040919\
Data File : VD061688.D
Acq On : 9 Apr 2019 21:25
Operator : VA/AP
Sample : K2316-05
Misc : 5.02a/5ml/MSVOA D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :

Quant Time: Apr 10 00:57:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.08 min Scan# 575

Delta R.T. 0.04 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25


Instrument :
MSVOA_D
ClientSampleId :

Tot Ion:168 Resp: 444631

Ion Ratio Lower Upper




168 100

99 54.8 46.8 70.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.95 (98.65 to 99.65): VDC



#16

Acetone

Concen: 54.348 ug/l

RT: 3.24 min Scan# 185

Delta R.T. 0.03 min

Lab File: VD061688.D


Acq: 9 Apr 2019 21:25

Tgt Ion: 43 Resp: 54755

Ion Ratio Lower Upper



43 100


58 22.9 19.7 29.5



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 57.95 (57.65 to 58.65): VDC





#17

Carbon Disulfide

Concen: 2.687 ug/l

RT: 3.41 min Scan# 202

Delta R.T. 0.03 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Instrument :

MSVOA_D


ClientSampleId :

Tot Ion: 76 Resp: 40476

Ion Ratio Lower Upper




76 100

78 9.5 7.4 11.2



Abundance Ion 75.85 (75.55 to 76.55): VDC

Ion 77.85 (77.55 to 78.55): VDC



#20

Methylene Chloride

Concen: 3.471 ug/l

RT: 3.85 min Scan# 247

Delta R.T. 0.05 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Tgt Ion: 84 Resp: 19841


Ion Ratio Lower Upper

84 100

49 123.4 93.0 139.4

51 35.4 30.4 45.6

86 58.8 51.0 76.4






Abundance Ion 83.95 (83.65 to 84.65): VDC

Ion 48.90 (48.60 to 49.60): VDC

Ion 50.90 (50.60 to 51.60): VDC

Ion 85.90 (85.60 to 86.60): VDC





#25

2-Butanone

Concen: 15.898 ug/l

RT: 6.10 min Scan# 475

Delta R.T. 0.03 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25


Instrument :
MSVOA_D
ClientSampled :

Tot Ion: 43 Resp: 18465

Ion Ratio Lower Upper




43 100

72 22.5 17.4 26.2



Abundance Ion 42.95 (42.65 to 43.65): VDC

Ion 72.00 (71.70 to 72.70): VDC



#33

1,2-Dichloroethane-d4

Concen: N.D. ug/l

RT: 7.49 min Scan# 616

Delta R.T. 0.03 min

Lab File: VD061688.D


Acq: 9 Apr 2019 21:25

Tgt Ion: 65 Resp: 184633

Ion Ratio Lower Upper



65 100


67 50.3 0.0 99.4



Abundance Ion 64.95 (64.65 to 65.65): VDC

Ion 66.90 (66.60 to 67.60): VDC





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.25 min Scan# 694

Delta R.T. 0.04 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Instrument :

MSVOA_D

ClientSampleId :


Tot Ion:114 Resp: 764376

Ion Ratio Lower Upper

114 100

63 15.2 0.0 31.2




88 15.8 0.0 31.8



Abundance Ion 114.00 (113.70 to 114.70): V

Ion 62.95 (62.65 to 63.65): V

Ion 87.90 (87.60 to 88.60): V



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 6.96 min Scan# 563

Delta R.T. 0.04 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25


Tgt Ion:113 Resp: 283408

Ion Ratio Lower Upper

113 100

111 102.3 79.3 118.9



192 16.8 12.3 18.5




Abundance Ion 112.90 (112.60 to 113.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 191.80 (191.50 to 192.50): V









#36
 1,1-Dichloropropene
 Concen: 4.746 ug/l
 RT: 7.08 min Scan# 575
 Delta R.T. -0.06 min
 Lab File: VD061688.D
 Acq: 9 Apr 2019 21:25




Instrument :
 MSVOA_D
 ClientSampled :


Tot Ion	75	Resp	31752
Ion Ratio	Lower	Upper	
75	100		
110	9.5	19.7	59.0#
77	0.0	24.8	37.2#



#37
 Ethyl Acetate
 Concen: 6.777 ug/l
 RT: 6.10 min Scan# 475
 Delta R.T. -0.07 min
 Lab File: VD061688.D
 Acq: 9 Apr 2019 21:25

Tgt Ion	43	Resp	18465
Ion Ratio	Lower	Upper	
43	100		
61	0.0	11.9	17.9#
70	0.0	6.9	10.3#









#38
Carbon Tetrachloride
Concen: 5.462 ug/l
RT: 7.08 min Scan# 575
Delta R.T. -0.02 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Instrument :
MSVOA_D
ClientSampleId :

Tot Ion:117 Resp: 42820
Ion Ratio Lower Upper
117 100
119 5.7 77.9 116.9#
121 0.0 24.7 37.1#




Abundance Ion 116.85 (116.55 to 117.55): V
Ion 118.85 (118.55 to 119.55): V
Ion 120.85 (120.55 to 121.55): V






#40
Benzene
Concen: 3.013 ug/l
RT: 7.49 min Scan# 616
Delta R.T. 0.03 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Tgt Ion: 78 Resp: 48622
Ion Ratio Lower Upper
78 100
77 24.0 19.5 29.3



Abundance Ion 77.95 (77.65 to 78.65): V
Ion 76.95 (76.65 to 77.65): V









#50
Toluene-d8
Concen: N.D. ug/l
RT: 10.44 min Scan# 916
Delta R.T. 0.03 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Instrument :
MSVOA_D
ClientSampleId :

Tot Ion: 98 Resp: 568091
Ion Ratio Lower Upper
98 100
100 69.0 55.4 83.0




Abundance Ion 98.05 (97.75 to 98.75): VD061644.D (-905) (-)
Ion 100.05 (99.75 to 100.75): VD061644.D (-905) (-)






#51
4-Methyl-2-Pentanone
Concen: 9.202 ug/l
RT: 10.39 min Scan# 911
Delta R.T. 0.09 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Tgt Ion: 43 Resp: 21521
Ion Ratio Lower Upper
43 100
58 24.1 29.8 44.8#



Abundance Ion 42.95 (42.65 to 43.65): VD061644.D (-894) (-)
Ion 57.95 (57.65 to 58.65): VD061644.D (-894) (-)





#56

Ethyl methacrylate

Concen: 1.019 ug/l

RT: 11.23 min Scan# 996

Delta R.T. 0.19 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Instrument :
MSVOA_D
ClientSampleId :


Tot Ion: 69 Resp: 3654

Ion Ratio Lower Upper




69 100

41 221.7 55.2 82.8#

39 91.9 28.6 43.0#



Abundance Ion 69.00 (68.70 to 69.70): VDC
Ion 41.05 (40.75 to 41.75): VDC
Ion 38.95 (38.65 to 39.65): VDC



#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 13.58 min Scan# 1235

Delta R.T. 0.03 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25


Tgt Ion: 95 Resp: 167297

Ion Ratio Lower Upper



95 100


174 97.6 0.0 164.6

176 95.7 0.0 165.0



Abundance Ion 95.00 (94.70 to 95.70): VDC
Ion 173.90 (173.60 to 174.60): VDC
Ion 175.90 (175.60 to 176.60): VDC









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 12.39 min Scan# 1114
Delta R.T. 0.03 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Instrument :
MSVOA_D
ClientSampleId :

Tot Ion:117 Resp: 470346
Ion Ratio Lower Upper
117 100
82 47.6 36.8 55.2
119 33.5 26.1 39.1




Abundance Ion 116.95 (116.65 to 117.65): V
Ion 82.05 (81.75 to 82.75): V
Ion 118.95 (118.65 to 119.65): V






#65
Chlorobenzene
Concen: 2.768 ug/l
RT: 12.43 min Scan# 1118
Delta R.T. 0.03 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Tgt Ion:112 Resp: 25815
Ion Ratio Lower Upper
112 100
114 39.5 25.6 38.4#



Abundance Ion 111.90 (111.60 to 112.60): V
Ion 114.00 (113.70 to 114.70): V









#67
Ethyl Benzene
Concen: 1.635 ug/l
RT: 12.55 min Scan# 1130
Delta R.T. 0.03 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25




Instrument :
MSVOA_D
ClientSampleId :


Tot Ion: 91 Resp: 23225
Ion Ratio Lower Upper
91 100
106 35.1 24.6 37.0



#68
m/p-Xylenes
Concen: 2.607 ug/l
RT: 12.68 min Scan# 1144
Delta R.T. 0.02 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Tgt Ion: 106 Resp: 14274
Ion Ratio Lower Upper
106 100
91 206.3 156.7 235.1









#69
o-Xylene
Concen: 2.145 ug/l
RT: 13.07 min Scan# 1183
Delta R.T. 0.03 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Instrument :
MSVOA_D
ClientSampled :

Tot Ion:106 Resp: 11339
Ion Ratio Lower Upper
106 100
91 209.4 107.3 321.9




Abundance Ion 106.05 (105.75 to 106.75): V
Ion 91.00 (90.70 to 91.70): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 14.53 min Scan# 1332
Delta R.T. 0.02 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Tgt Ion:152 Resp: 114800
Ion Ratio Lower Upper
152 100
115 59.2 26.4 79.2
150 161.2 0.0 320.6



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 115.00 (114.70 to 115.70): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 13.244 ug/l

RT: 13.42 min Scan# 1219

Delta R.T. 0.03 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Instrument :

MSVOA_D


ClientSampleId :

Tot Ion:105 Resp: 101303

Ion Ratio Lower Upper




105 100

120 24.5 12.2 36.4



Abundance Ion 105.05 (104.75 to 105.75): V

Ion 120.05 (119.75 to 120.75): V



#74

N-ethyl acetate

Concen: 7.615 ug/l

RT: 13.24 min Scan# 1200

Delta R.T. -0.02 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Tgt Ion: 43 Resp: 15827


Ion Ratio Lower Upper

43 100

70 11.0 25.4 38.2#

55 43.2 15.0 22.4#

61 0.0 20.9 31.3#






Abundance Ion 42.95 (42.65 to 43.65): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 60.95 (60.65 to 61.65): V









#76
 1,2,3-Trichloropropane
 Concen: 39.883 ug/l
 RT: 13.58 min Scan# 1235
 Delta R.T. -0.15 min
 Lab File: VD061688.D
 Acq: 9 Apr 2019 21:25




Instrument :
 MSVOA_D
 ClientSampled :


Tot Ion: 75 Resp: 62292
 Ion Ratio Lower Upper
 75 100
 77 2.3 15.7 47.1#



#78
 n-propylbenzene
 Concen: 2.769 ug/l
 RT: 13.78 min Scan# 1255
 Delta R.T. 0.01 min
 Lab File: VD061688.D
 Acq: 9 Apr 2019 21:25

Tgt Ion: 91 Resp: 25332
 Ion Ratio Lower Upper
 91 100
 120 0.0 10.8 32.4#









#79
2-Chlorotoluene
Concen: 14.572 ug/l
RT: 13.86 min Scan# 1263
Delta R.T. 0.02 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25




Instrument :
MSVOA_D
ClientSampleId :


Tot Ion: 91 Resp: 76247
Ion Ratio Lower Upper
91 100
126 35.0 20.0 59.9



#80
1,3,5-Trimethylbenzene
Concen: 1.797 ug/l
RT: 13.95 min Scan# 1273
Delta R.T. 0.03 min
Lab File: VD061688.D
Acq: 9 Apr 2019 21:25

Tgt Ion:105 Resp: 10910
Ion Ratio Lower Upper
105 100
120 53.3 24.4 73.2





#81

trans-1,4-Dichloro-2-butene

Concen: 131.834 ug/l

RT: 13.58 min Scan# 1235

Delta R.T. 0.10 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Instrument :

MSVOA_D

ClientSampleId :


Tot Ion: 75 Resp: 62483

Ion Ratio Lower Upper




75 100

53 1.3 60.2 90.2#

89 0.0 34.2 51.2#



Abundance Ion 75.00 (74.70 to 75.70): VDC
Ion 53.00 (52.70 to 53.70): VDC
Ion 89.00 (88.70 to 89.70): VDC



#83

tert-Butylbenzene

Concen: 3.600 ug/l

RT: 14.21 min Scan# 1299

Delta R.T. 0.03 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25


Tgt Ion: 119 Resp: 26351

Ion Ratio Lower Upper



119 100


91 64.1 33.9 101.6

134 23.6 11.6 34.8



Abundance Ion 119.05 (118.75 to 119.75): VDC
Ion 91.00 (90.70 to 91.70): VDC
Ion 134.00 (133.70 to 134.70): VDC









#84
 1,2,4-Trimethylbenzene
 Concen: 3.427 ug/l
 RT: 14.26 min Scan# 1304
 Delta R.T. 0.03 min
 Lab File: VD061688.D
 Acq: 9 Apr 2019 21:25




Instrument :
 MSVOA_D
 ClientSampleId :


Tot Ion:105 Resp: 20999
 Ion Ratio Lower Upper
 105 100
 120 42.6 22.6 67.8



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 4.307 ug/l
 RT: 15.35 min Scan# 1415
 Delta R.T. -0.02 min
 Lab File: VD061688.D
 Acq: 9 Apr 2019 21:25

Tgt Ion: 75 Resp: 822
 Ion Ratio Lower Upper
 75 100
 155 39.7 45.6 136.7#
 157 0.0 61.4 184.2#





#95

Naphthalene

Concen: 6.738 ug/l

RT: 16.14 min Scan# 1495

Delta R.T. 0.02 min

Lab File: VD061688.D

Acq: 9 Apr 2019 21:25

Instrument :

MSVOA_D

ClientSampleId :


Tot Ion:128 Resp: 17107

Ion Ratio Lower Upper

128 100

127 15.0 9.4 14.2#

129 11.4 8.8 13.2



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

