

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040919\
 Data File : VD061683.D
 Acq On : 9 Apr 2019 19:06
 Operator : VA/AP
 Sample : K2294-06RE
 Misc : 2.99g/5ml/MSVOA D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-14-D2RE

Quant Time: Apr 10 02:12:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.09	168	410547	50.00	ug/l	0.05
34) 1,4-Difluorobenzene	8.25	114	565962	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.39	117	322678	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.53	152	62171	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.49	65	167019	53.67	ug/l	0.04
Spiked Amount 50.000			Recovery	=	107.34%	
35) Dibromofluoromethane	6.97	113	274040	69.89	ug/l	0.05
Spiked Amount 50.000			Recovery	=	139.78%	
50) Toluene-d8	10.44	98	529139	53.72	ug/l	0.03
Spiked Amount 50.000			Recovery	=	107.44%	
62) 4-Bromofluorobenzene	13.57	95	112192	29.69	ug/l	0.01
Spiked Amount 50.000			Recovery	=	59.38%	
Target Compounds						
16) Acetone	3.25	43	29496	31.707	ug/l	98
20) Methylene Chloride	3.86	84	11018	2.087	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040919\
Data File : VD061683.D
Acq On : 9 Apr 2019 19:06
Operator : VA/AP
Sample : K2294-06RE
Misc : 2.99g/5ml/MSVOA D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-14-D2RE

Quant Time: Apr 10 02:12:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

