

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040919\
 Data File : VD061696.D
 Acq On : 10 Apr 2019 2:58
 Operator : VA/AP
 Sample : VD0409SBL02
 Misc : 5.00µl/5ml/MSVOA D/SOIL
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0409SBL02

Quant Time: Apr 10 05:18:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.07	168	637484	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.25	114	1099255	50.00	ug/l	0.04
63) Chlorobenzene-d5	12.38	117	882814	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.53	152	432791	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	247245	51.17	ug/l	0.03
Spiked Amount 50.000			Recovery	=	102.34%	
35) Dibromofluoromethane	6.96	113	374880	49.23	ug/l	0.04
Spiked Amount 50.000			Recovery	=	98.46%	
50) Toluene-d8	10.43	98	888058	46.42	ug/l	0.03
Spiked Amount 50.000			Recovery	=	92.84%	
62) 4-Bromofluorobenzene	13.57	95	341529	50.99	ug/l	0.02
Spiked Amount 50.000			Recovery	=	101.98%	
Target Compounds						
20) Methylene Chloride	3.85	84	25645	3.129	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040919\
Data File : VD061696.D
Acq On : 10 Apr 2019 2:58
Operator : VA/AP
Sample : VD0409SBL02
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0409SBL02

Quant Time: Apr 10 05:18:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

