

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD040919\
 Data File : VD061702.D
 Acq On : 10 Apr 2019 6:12
 Operator : VA/AP
 Sample : K2338-10
 Misc : 6.07g/5ml/MSVOA D/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-2-DEEP

Quant Time: Apr 10 06:48:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
 Quant Title : SW846 8260
 QLast Update : Mon Apr 08 10:04:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.08	168	115793	50.00	ug/l	0.04
34) 1,4-Difluorobenzene	8.24	114	233296	50.00	ug/l	0.03
63) Chlorobenzene-d5	12.38	117	205975	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	14.53	152	107617	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.48	65	105393	120.08	ug/l	0.03
Spiked Amount 50.000			Recovery	=	240.16%	
35) Dibromofluoromethane	6.95	113	119613	74.01	ug/l	0.03
Spiked Amount 50.000			Recovery	=	148.02%	
50) Toluene-d8	10.43	98	243806	60.04	ug/l	0.03
Spiked Amount 50.000			Recovery	=	120.08%	
62) 4-Bromofluorobenzene	13.56	95	109720	92.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	185.20%	
Target Compounds						
16) Acetone	3.24	43	25957	98.931	ug/l	94
20) Methylene Chloride	3.85	84	32385	21.752	ug/l	94
52) Toluene	10.53	92	4826	1.569	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD040919\
Data File : VD061702.D
Acq On : 10 Apr 2019 6:12
Operator : VA/AP
Sample : K2338-10
Misc : 6.07g/5ml/MSVOA D/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-2-DEEP

Quant Time: Apr 10 06:48:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D040619S.M
Quant Title : SW846 8260
QLast Update : Mon Apr 08 10:04:03 2019
Response via : Initial Calibration

