

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040921\
 Data File : VD068818.D
 Acq On : 09 Apr 2021 18:11
 Operator : VA/SY
 Sample : M1979-04
 Misc : 5.32G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-4

Quant Time: Apr 10 00:45:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 08 03:50:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	473411	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	830006	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	842741	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.573	152	437777	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	201608	40.30	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.60%
35) Dibromofluoromethane	7.914	113	253095	46.88	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.76%
50) Toluene-d8	10.338	98	986486	48.44	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.88%
62) 4-Bromofluorobenzene	12.632	95	395109	58.41	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD040921\
Data File : VD068818.D
Acq On : 09 Apr 2021 18:11
Operator : VA/SY
Sample : M1979-04
Misc : 5.32G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-4

Quant Time: Apr 10 00:45:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D040721S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 08 03:50:31 2021
Response via : Initial Calibration

